

Table S 4: Validation study results (ambulatory validation study), for mean calculation of inflation and deflation measurements.

	<i>Pass</i>	<i>Achieved</i>	
	<i>requirement</i>	<i>SBP</i>	<i>DBP</i>
Criterion 1 (106 BP pairs)			
Mean BP difference (mmHg)	≤ 5	-1.0	2.8
SD (mmHg)	≤ 8	7.0	6.2
		<i>Pass</i>	<i>Pass</i>
Result		<i>Pass</i>	